

CERTIFICATE OF ANALYSIS

16/100: Pre-packed Column

Resin: XXXXXXXXXXXX
Resin lot#: 089001CO


Theoretical Plate Number and Peak Asymmetry test

Tracer: 0.2 ml of 1M NaCl
Flowrate: 5 ml/min
Mobile phase: 0.15M NaCl

Test Results

Test	Theoretical plate number (N/m)	Peak asymmetry	Bed height (cm)	Delta column pressure at 5 ml/min (MPa)	Packing batch
Column #1	5041	1.1	10	0.15	D28042301

Manufactured in UK under quality system management.

A handwritten signature in black ink, appearing to be "Oliver Watson".

Signature

Oliver Watson
Print name

10-May-23 / 8.41
Date/Time

Production Scientist
Title